

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | |
|--|--|
| DataMaster cdm S/N <u>130306</u> ✓ | |
| Supervisor/Operator Performing the Verification Procedure: | |
| Name <u>STEVE DUNN</u> | ID# <u>4106</u> ✓ Date <u>3-4-11</u> ✓ |
| A Agency <u>APD</u> | Phone # <u>786-8900</u> |
| Instrument Location <u>4th AVE SUB</u> | |
| B Alco S/N <u>X124669</u> ✓ Target Value <u>.077</u> ✓ High Pressure <u>1000 PSI</u> | |
| Alco Test Values <u>.073</u> ✓ <u>.073</u> ✓ | |
| 1 st Alco 2 nd Alco | |
| Signature <u>[Signature]</u> | |
| (OVER) | |

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAR, 2011 AN

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130306

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

S. DUNN
#1361

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

MARCH 04, 2011

OPERATOR'S NAME:

DUNN/S

OPERATOR'S NUMBER: 4106

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

U

O.L. #: U

DEPT/AGENCY: ANCI

CASE/REPORT: U

TEST TYPE: U

ALCO TARGET VALUE: .077

ALCO S/N: X124669

--- BREATH ANALYSIS ---

| | | |
|----------------------------|----------|-------|
| .077 ADJUSTED FOR 30.16 in | | |
| ALCO TARGET | .077 | 10:35 |
| BLANK TEST | .000 | 10:37 |
| INTERNAL STANDARD | VERIFIED | 10:37 |
| ALCO TV 30.16 in | .073 | 10:37 |
| BLANK TEST | .000 | 10:38 |
| SUBJECT SAMPLE | .000 | 10:38 |
| BLANK TEST | .000 | 10:39 |
| ALCO TV 30.16 in | .073 | 10:40 |
| BLANK TEST | .000 | 10:40 |

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

MARCH 04, 2011

TIME 10:28

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 41c |
| BAROMETER: | 30.16 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

"#%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKL"NOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~||